



Memo

ATTENTION	To the Manager ,JAWAS MEDICAL Doha - Qatar	DATE	28-1-2020
FROM	Mr. Ibrahim Falamerzi, Customer Service Manager – Supply Chain Management	<i>me 28/1/2020</i>	
SUBJECT	Evaluation Report for Prequalification Sample		
CC			
<input type="checkbox"/> URGENT	<input type="checkbox"/> CONFIDENTIAL	<input type="checkbox"/> PLEASE PROCESS	<input type="checkbox"/> FYI

We would like to inform you the evaluation feedback for your items.

Manufacture	Country	Cat #	Description	Positive points	Negative points	Recommendations.
I - SENS	KOREA	I-SENS	CARE SENS / BLOOD GULCOSE MONTORING SYSTEM / BLOOD GULCOSE TEST	<ul style="list-style-type: none">- Easy to use / big and clear result numbers- No need coding- High operating range of temperature and humidity- Device is user friendly , light weighed , easy to use and transport- Strip quality – not affected easily by environmental factors like room temperature- Minimum volume of blood drop and acceptable for patients with all age groups- Requires less time to produce glucose result and is reliable	<ul style="list-style-type: none">- Cannot be used for new born- Default unit measurement (UoM) is mg/dl which cant be changed	ACCETABLE

Note:

Prequalification evaluation not for purchase, should participate in the tender